



Product Service

Attestation of Conformity

No. E8A 111906 0004 Rev. 00

Holder of Attestation: **Gas DNA Co., Ltd.**
101, Bukhang-ro 193beon-gil, Seo-gu
Incheon 22856
REPUBLIC OF KOREA

Name of Object: **Gas Detection Systems
(Smart Gas Detector)**

Model(s): **DA-600S**

Description of Object: Rated input voltage: 24 V d.c.

Tested according to: EN 50270:2015
EN 61000-6-2:2005
EN IEC 61000-6-2:2019
EN 61000-6-4:2007/A1:2011
EN IEC 61000-6-4:2019

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.: CTK-2025-01642

Date, 2025-06-27

(Yong-Tae Yu)

Page 1 of 1

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.