



No. 2010-408

Safety Certificate

- Manufacturer Gas DNA Co., Ltd.
- Address 101, Bukhang-ro 193beon-gil, Seo-gu, Incheon,
Republic of KOREA

The following product fabricated by the above manufacturer has been certified for the use of the Safety Certification Mark as the product was assessed and tested in accordance with the article 34 of the Occupational Safety & Health Act and the article 58 (4) 4 of the Enforcement Rule thereof.

- Product Gas Detector
- Type (Model) DA-100
- Rating Max. DC 24 V, 120 mA
- Type of Protection Ex d IIC T5
- Certification No. 10-GA2BO-408
- Safety Certification Standards Article 34 of Industrial Safety & Health Law
- Certification Condition The products manufactured in above address only.
- Date of Issue June 3rd, 2011

Approved for issue on behalf of the president : Weon-Don, Joo
Position : General Manager

Signature



KOREA GAS SAFETY CORPORATION

