



No. 2015-0343

Safety Certificate

- Manufacturer Gas DNA Co., Ltd.
- Address 101, Bukhang-ro 193beon-gil, Seo-gu, Incheon, Republic of KOREA

The following product fabricated by the above manufacturer has been certified for the use of the Safety Certification Mark as the product was assessed and tested in accordance with the article 34 of the Occupational Safety & Health Act and the article 58 (4) 4 of the Enforcement Rule thereof.

- Product Gas Detector
- Type (Model) DA-800
- Rating DC 20 ~ 30 V, 500 mA
- Type of Protection Ex d IIB+H2 T6
- Certification No. 15-GA2BO-0343
- Safety Certification Standards Notification of Protective device Mandatory Safety Certification (Notice of Ministry of Employment and of Labor No. 2013-54)
- Certification Condition
 1. Tamb = - 20°C ~ + 60°C
 2. Each enclosure shall be subjected to strength test of 10.3 bar
 3. Refer to the instruction manual before operation.
- Date of Issue May 6th, 2015

Approved for issue on behalf of the president : Weon-Don, Joo
Position : General Manager

Signature



KOREA GAS SAFETY CORPORATION

